

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026368.D
 Acq On : 06 Jan 2022 12:51
 Operator : JC/MD
 Sample : VX0106WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0106WBL01

Quant Time: Jan 07 04:11:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	155164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	259722	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92269	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	110686	49.308	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.620%
35) Dibromofluoromethane	5.391	113	83595	49.507	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.020%
50) Toluene-d8	8.653	98	299656	47.710	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.420%
62) 4-Bromofluorobenzene	11.079	95	107875	48.248	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026368.D
Acq On : 06 Jan 2022 12:51
Operator : JC/MD
Sample : VX0106WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0106WBL01

Quant Time: Jan 07 04:11:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

