

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026372.D
 Acq On : 06 Jan 2022 14:44
 Operator : JC/MD
 Sample : N1019-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-7DL

Quant Time: Jan 07 04:13:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151570	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	251881	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95048	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	107344	48.954	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.900%
35) Dibromofluoromethane	5.391	113	81140	49.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.100%
50) Toluene-d8	8.653	98	292508	48.022	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.040%
62) 4-Bromofluorobenzene	11.079	95	108311	49.951	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.900%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2698	2.271	ug/l	99
31) Cyclohexane	5.477	56	28880	9.248	ug/l	92
39) Methylcyclohexane	7.385	83	7040	2.363	ug/l	88
40) Benzene	6.044	78	53915	7.152	ug/l	96
52) Toluene	8.720	92	40168	8.656	ug/l	99
67) Ethyl Benzene	10.195	91	43708	5.172	ug/l	97
68) m/p-Xylenes	10.299	106	287226	91.022	ug/l	93
69) o-Xylene	10.640	106	3248	1.052	ug/l	92
73) Isopropylbenzene	10.963	105	57552	7.790	ug/l	98
78) n-propylbenzene	11.305	91	112670	13.264	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	32660	5.327	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	2533	0.413	ug/l	99
85) sec-Butylbenzene	11.890	105	1738	0.234	ug/l #	56
89) n-Butylbenzene	12.335	91	2638	0.479	ug/l #	46
95) Naphthalene	13.774	128	77580	10.561	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026372.D
Acq On : 06 Jan 2022 14:44
Operator : JC/MD
Sample : N1019-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-7DL

Quant Time: Jan 07 04:13:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

