

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026373.D
 Acq On : 06 Jan 2022 15:07
 Operator : JC/MD
 Sample : N1019-05DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-26DL

Quant Time: Jan 07 04:14:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	153286	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	253905	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216691	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97131	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109355	49.312	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	98.620%	
35) Dibromofluoromethane	5.391	113	81867	49.595	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.180%	
50) Toluene-d8	8.653	98	294212	47.917	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	95.840%	
62) 4-Bromofluorobenzene	11.079	95	108272	49.535	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.060%	
Target Compounds						
					Qvalue	
31) Cyclohexane	5.477	56	5084	1.610	ug/l	93
39) Methylcyclohexane	7.385	83	993	0.331	ug/l #	84
40) Benzene	6.044	78	49449	6.508	ug/l	100
52) Toluene	8.720	92	66510	14.218	ug/l	99
67) Ethyl Benzene	10.195	91	296666	34.855	ug/l	98
68) m/p-Xylenes	10.305	106	273632	86.101	ug/l	93
69) o-Xylene	10.640	106	75905	24.411	ug/l	91
73) Isopropylbenzene	10.963	105	11942	1.582	ug/l	99
78) n-propylbenzene	11.305	91	23080	2.659	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	27514	4.391	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	118016	18.851	ug/l	98
95) Naphthalene	13.780	128	37994	5.061	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026373.D
Acq On : 06 Jan 2022 15:07
Operator : JC/MD
Sample : N1019-05DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-26DL

Quant Time: Jan 07 04:14:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

