

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026374.D
 Acq On : 06 Jan 2022 15:30
 Operator : JC/MD
 Sample : N1019-06DL 40X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7DL

Quant Time: Jan 07 04:14:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150571	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	254148	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220921	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109410	50.227	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 100.460%			
35) Dibromofluoromethane	5.391	113	81209	49.149	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.300%			
50) Toluene-d8	8.653	98	294064	47.847	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 95.700%			
62) 4-Bromofluorobenzene	11.079	95	110995	50.732	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 101.460%			
Target Compounds					Qvalue	
31) Cyclohexane	5.476	56	3234	1.043	ug/l	95
39) Methylcyclohexane	7.384	83	1220	0.406	ug/l	99
40) Benzene	6.043	78	30034	3.949	ug/l	91
52) Toluene	8.720	92	20764	4.435	ug/l	97
67) Ethyl Benzene	10.195	91	125625	14.477	ug/l	98
68) m/p-Xylenes	10.299	106	301351	93.008	ug/l	92
69) o-Xylene	10.640	106	68965	21.754	ug/l	92
73) Isopropylbenzene	10.963	105	4961	0.652	ug/l	99
78) n-propylbenzene	11.305	91	8567	0.979	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	33976	5.381	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	123601	19.590	ug/l	98
95) Naphthalene	13.780	128	30209	3.993	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026374.D
Acq On : 06 Jan 2022 15:30
Operator : JC/MD
Sample : N1019-06DL 40X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7DL

Quant Time: Jan 07 04:14:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

