

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026375.D
 Acq On : 06 Jan 2022 15:53
 Operator : JC/MD
 Sample : N1020-07DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7ADL

Quant Time: Jan 07 04:14:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	152202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	251922	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216899	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90887	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108202	49.140	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	98.280%		
35) Dibromofluoromethane	5.391	113	81166	49.557	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.120%		
50) Toluene-d8	8.653	98	292307	47.981	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	95.960%		
62) 4-Bromofluorobenzene	11.079	95	105625	48.704	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.400%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	3069	2.572	ug/l	95
31) Cyclohexane	5.477	56	9917	3.163	ug/l	96
39) Methylcyclohexane	7.385	83	1190	0.399	ug/l	95
40) Benzene	6.044	78	398047	52.796	ug/l	99
52) Toluene	8.720	92	43828	9.443	ug/l	100
67) Ethyl Benzene	10.195	91	41348	4.853	ug/l	96
68) m/p-Xylenes	10.305	106	68378	21.495	ug/l	92
69) o-Xylene	10.640	106	5497	1.766	ug/l	98
73) Isopropylbenzene	10.964	105	14052	1.989	ug/l	98
78) n-propylbenzene	11.305	91	23103	2.844	ug/l	96
80) 1,3,5-Trimethylbenzene	11.451	105	5979	1.020	ug/l	93
84) 1,2,4-Trimethylbenzene	11.750	105	18660	3.185	ug/l	97
95) Naphthalene	13.774	128	43756	6.229	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026375.D
Acq On : 06 Jan 2022 15:53
Operator : JC/MD
Sample : N1020-07DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7ADL

Quant Time: Jan 07 04:14:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

