

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026377.D
 Acq On : 06 Jan 2022 16:40
 Operator : JC/MD
 Sample : N1020-10DL 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-24DL

Quant Time: Jan 07 04:15:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	247969	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	107481	50.322	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.640%		
35) Dibromofluoromethane	5.391	113	79512	49.321	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 98.640%		
50) Toluene-d8	8.653	98	285707	47.646	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 95.300%		
62) 4-Bromofluorobenzene	11.079	95	105196	49.280	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 98.560%		
Target Compounds						
31) Cyclohexane	5.489	56	3415	1.123	ug/l	Qvalue # 89
40) Benzene	6.044	78	699502	94.260	ug/l	98
52) Toluene	8.720	92	16234	3.553	ug/l	100
67) Ethyl Benzene	10.195	91	274623	33.250	ug/l	97
68) m/p-Xylenes	10.299	106	341381	110.700	ug/l	93
69) o-Xylene	10.640	106	40186	13.318	ug/l	91
73) Isopropylbenzene	10.964	105	6110	0.840	ug/l	95
78) n-propylbenzene	11.305	91	12098	1.446	ug/l	95
80) 1,3,5-Trimethylbenzene	11.451	105	19811	3.281	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	82675	13.705	ug/l	97
95) Naphthalene	13.780	128	29536	4.083	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026377.D
Acq On : 06 Jan 2022 16:40
Operator : JC/MD
Sample : N1020-10DL 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-24DL

Quant Time: Jan 07 04:15:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

