

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026380.D
 Acq On : 06 Jan 2022 17:49
 Operator : JC/MD
 Sample : N1020-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-36

Quant Time: Jan 07 04:17:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	248268	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209761	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103890	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.977	65	103734	48.163	ug/l	0.01
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.320%		
35) Dibromofluoromethane	5.391	113	78491	48.629	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.260%		
50) Toluene-d8	8.653	98	282662	47.081	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.160%		
62) 4-Bromofluorobenzene	11.085	95	108646	50.834	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.660%		
Target Compounds					Qvalue	
16) Acetone	2.386	43	9289	7.960	ug/l	91
22) Diisopropyl ether	3.764	45	52654	8.468	ug/l #	75
31) Cyclohexane	5.483	56	180835	58.957	ug/l	96
37) Ethyl Acetate	4.733	43	5657	1.874	ug/l	97
39) Methylcyclohexane	7.385	83	31096	10.589	ug/l	92
40) Benzene	6.086	78	23162677	3117.484	ug/l #	78
52) Toluene	8.720	92	805808	176.170	ug/l	98
67) Ethyl Benzene	10.208	91	9489006	1151.673	ug/l #	82
68) m/p-Xylenes	10.336	106	10671127	3468.715	ug/l	77
69) o-Xylene	10.653	106	1836795	610.221	ug/l	98
73) Isopropylbenzene	10.963	105	318161	39.399	ug/l	98
78) n-propylbenzene	11.305	91	656387	70.695	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	983724	146.791	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	3682480	549.940	ug/l	97
95) Naphthalene	13.780	128	1715705	213.680	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026380.D
Acq On : 06 Jan 2022 17:49
Operator : JC/MD
Sample : N1020-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-36

Quant Time: Jan 07 04:17:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

