

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026382.D
 Acq On : 06 Jan 2022 18:35
 Operator : JC/MD
 Sample : N1020-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Jan 07 04:17:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	153654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	257270	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218711	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92187	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	107826	48.506	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	97.020%	
35) Dibromofluoromethane	5.391	113	81898	48.965	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.920%	
50) Toluene-d8	8.653	98	295105	47.434	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	94.860%	
62) 4-Bromofluorobenzene	11.079	95	107129	48.371	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	96.740%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	16789	13.940	ug/l	97
40) Benzene	6.044	78	6787	0.882	ug/l	100
67) Ethyl Benzene	10.195	91	4911	0.572	ug/l	96
68) m/p-Xylenes	10.299	106	6886	2.147	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	1849	0.311	ug/l	99
95) Naphthalene	13.780	128	1897	0.266	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026382.D
Acq On : 06 Jan 2022 18:35
Operator : JC/MD
Sample : N1020-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Jan 07 04:17:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

