

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026385.D
 Acq On : 06 Jan 2022 19:45
 Operator : JC/MD
 Sample : N1020-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-01-04

Quant Time: Jan 07 04:18:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 30 08:02:21 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	252785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90599	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108064	49.486	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.980%
35) Dibromofluoromethane	5.391	113	80438	48.945	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	8.653	98	290174	47.469	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.940%
62) 4-Bromofluorobenzene	11.079	95	106572	48.973	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.940%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1517	1.282	ug/l #	85
68) m/p-Xylenes	10.299	106	1918	0.609	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
Data File : VX026385.D
Acq On : 06 Jan 2022 19:45
Operator : JC/MD
Sample : N1020-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-01-04

Quant Time: Jan 07 04:18:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 30 08:02:21 2021
Response via : Initial Calibration

