


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010622\
 Data File : VX026366.D
 Acq On : 06 Jan 2022 10:19
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X123021W.M
 Title : SW846 8260
 Last Update : Thu Dec 30 08:02:21 2021



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	6041	PASS
75	95	30	60	53.7	16705	PASS
95	95	100	100	100.0	31096	PASS
96	95	5	9	6.7	2078	PASS
173	174	0.00	2	1.1	225	PASS
174	95	50	100	68.7	21375	PASS
175	174	5	9	8.0	1705	PASS
176	174	95	101	95.4	20388	PASS
177	176	5	9	6.7	1358	PASS