

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
 Data File : VX033620.D
 Acq On : 06 Jan 2023 10:54
 Operator : JC/MD
 Sample : VX0106WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0106WBL01

Quant Time: Jan 07 04:20:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	167912	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	240297	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209059	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98687	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88959	46.931	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.860%
35) Dibromofluoromethane	5.385	113	74406	47.109	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.220%
50) Toluene-d8	8.647	98	273418	48.432	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.860%
62) 4-Bromofluorobenzene	11.079	95	96275	48.113	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
Data File : VX033620.D
Acq On : 06 Jan 2023 10:54
Operator : JC/MD
Sample : VX0106WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0106WBL01

Quant Time: Jan 07 04:20:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

