

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
 Data File : VX033623.D
 Acq On : 06 Jan 2023 12:19
 Operator : JC/MD
 Sample : PB050057TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB050057TB

Quant Time: Jan 07 04:22:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	183426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	261686	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236161	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96854	46.775	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.540%		
35) Dibromofluoromethane	5.385	113	81341	47.290	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.580%		
50) Toluene-d8	8.647	98	305554	49.701	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.400%		
62) 4-Bromofluorobenzene	11.079	95	110128	50.537	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.080%		
Target Compounds						
16) Acetone	2.392	43	23297	20.948	ug/l	98
25) 2-Butanone	4.574	43	7016	5.189	ug/l	95
37) Ethyl Acetate	4.715	43	24965	7.880	ug/l	100
43) Isopropyl Acetate	6.336	43	104440	26.218	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
Data File : VX033623.D
Acq On : 06 Jan 2023 12:19
Operator : JC/MD
Sample : PB050057TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB050057TB

Quant Time: Jan 07 04:22:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

