

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
 Data File : VX033624.D
 Acq On : 06 Jan 2023 12:42
 Operator : JC/MD
 Sample : PB050057ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB050057ZHE#01

Quant Time: Jan 07 04:22:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	176027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	254875	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	107905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	96120	48.372	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.740%
35) Dibromofluoromethane	5.385	113	79393	47.391	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	8.647	98	295832	49.406	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.820%
62) 4-Bromofluorobenzene	11.079	95	105947	49.918	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.840%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	24064	23.613	ug/l	96
25) 2-Butanone	4.574	43	7178	5.532	ug/l #	89
37) Ethyl Acetate	4.715	43	25183	8.241	ug/l	99
43) Isopropyl Acetate	6.336	43	108766	28.034	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
Data File : VX033624.D
Acq On : 06 Jan 2023 12:42
Operator : JC/MD
Sample : PB050057ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB050057ZHE#01

Quant Time: Jan 07 04:22:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

