

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
 Data File : VX033628.D
 Acq On : 06 Jan 2023 14:15
 Operator : JC/MD
 Sample : PB050057ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB050057ZHE#05

Quant Time: Jan 07 04:24:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129248	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	247910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106733	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72449	49.655	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	99.320%	
35) Dibromofluoromethane	5.391	113	55012	33.760	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	67.520%#	
50) Toluene-d8	8.647	98	288580	49.548	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	99.100%	
62) 4-Bromofluorobenzene	11.079	95	103489	50.129	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	100.260%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	25772	40.845	ug/l	95
18) Methyl Acetate	2.709	43	2276	1.125	ug/l	96
20) Methylene Chloride	2.788	84	5515	2.389	ug/l	93
25) 2-Butanone	4.574	43	7403	7.771	ug/l	96
37) Ethyl Acetate	4.714	43	24384	8.193	ug/l	100
43) Isopropyl Acetate	6.336	43	105367	27.921	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
Data File : VX033628.D
Acq On : 06 Jan 2023 14:15
Operator : JC/MD
Sample : PB050057ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB050057ZHE#05

Quant Time: Jan 07 04:24:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

