

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
 Data File : VX033632.D
 Acq On : 06 Jan 2023 15:48
 Operator : JC/MD
 Sample : PB050057ZHE#09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB050057ZHE#09

Quant Time: Jan 07 04:25:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	159881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	231267	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206666	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98996	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88298	48.923	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.840%
35) Dibromofluoromethane	5.385	113	71421	46.985	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.960%
50) Toluene-d8	8.647	98	270146	49.721	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.440%
62) 4-Bromofluorobenzene	11.079	95	95428	49.551	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.100%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	24974	28.972	ug/l	100
25) 2-Butanone	4.574	43	7238	6.142	ug/l	97
37) Ethyl Acetate	4.720	43	25836	9.608	ug/l	99
43) Isopropyl Acetate	6.342	43	105320	29.916	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
Data File : VX033632.D
Acq On : 06 Jan 2023 15:48
Operator : JC/MD
Sample : PB050057ZHE#09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB050057ZHE#09

Quant Time: Jan 07 04:25:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

