

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
 Data File : VX033633.D
 Acq On : 06 Jan 2023 16:11
 Operator : JC/MD
 Sample : PB050057ZHE#10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB050057ZHE#10

Quant Time: Jan 07 04:26:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	168777	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246906	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78293	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94013	49.344	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.680%
35) Dibromofluoromethane	5.391	113	77327	47.648	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.300%
50) Toluene-d8	8.647	98	289434	49.897	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.800%
62) 4-Bromofluorobenzene	11.079	95	103301	50.242	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.480%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	28016	31.663	ug/l	93
20) Methylene Chloride	2.788	84	5197	1.083	ug/l	91
25) 2-Butanone	4.574	43	7362	5.918	ug/l	100
37) Ethyl Acetate	4.721	43	24562	8.312	ug/l	99
43) Isopropyl Acetate	6.342	43	106865	28.433	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
Data File : VX033633.D
Acq On : 06 Jan 2023 16:11
Operator : JC/MD
Sample : PB050057ZHE#10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB050057ZHE#10

Quant Time: Jan 07 04:26:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

