

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
 Data File : VX033636.D
 Acq On : 06 Jan 2023 17:20
 Operator : JC/MD
 Sample : PB050057ZHE#13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB050057ZHE#13

Quant Time: Jan 07 04:27:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	167928	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	247326	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225343	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94781	49.998	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 100.000%		
35) Dibromofluoromethane	5.391	113	77460	47.649	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.300%		
50) Toluene-d8	8.647	98	289634	49.847	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.700%		
62) 4-Bromofluorobenzene	11.079	95	103098	50.058	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 100.120%		
Target Compounds						
						Qvalue
16) Acetone	2.392	43	24443	26.045	ug/l	99
20) Methylene Chloride	2.788	84	5890	1.554	ug/l	98
25) 2-Butanone	4.574	43	7049	5.695	ug/l	99
37) Ethyl Acetate	4.715	43	24559	8.293	ug/l	99
43) Isopropyl Acetate	6.336	43	104933	27.871	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
Data File : VX033636.D
Acq On : 06 Jan 2023 17:20
Operator : JC/MD
Sample : PB050057ZHE#13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB050057ZHE#13

Quant Time: Jan 07 04:27:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

