

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
 Data File : VX033637.D
 Acq On : 06 Jan 2023 17:44
 Operator : JC/MD
 Sample : PB050057ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB050057ZHE#14

Quant Time: Jan 07 04:27:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jan 05 04:43:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	163517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238499	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91951	49.814	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.620%
35) Dibromofluoromethane	5.385	113	75896	48.415	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.820%
50) Toluene-d8	8.647	98	284357	50.750	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.500%
62) 4-Bromofluorobenzene	11.079	95	100766	50.737	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.480%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	34413	43.884	ug/l	100
20) Methylene Chloride	2.788	84	6912	2.345	ug/l	98
25) 2-Butanone	4.574	43	7062	5.859	ug/l #	88
37) Ethyl Acetate	4.715	43	23247	8.100	ug/l	98
43) Isopropyl Acetate	6.336	43	101773	28.032	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010623\
Data File : VX033637.D
Acq On : 06 Jan 2023 17:44
Operator : JC/MD
Sample : PB050057ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB050057ZHE#14

Quant Time: Jan 07 04:27:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010423W.M
Quant Title : SW846 8260
QLast Update : Thu Jan 05 04:43:25 2023
Response via : Initial Calibration

