

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010625\
 Data File : VX044584.D
 Acq On : 06 Jan 2025 10:44
 Operator : JC/MD
 Sample : VX0106MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0106MBL01

Quant Time: Jan 06 16:02:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	131761	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	231608	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	210109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	96904	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	92725	40.137	ug/l	-0.01
Spiked Amount	50.000	Range 74 - 125	Recovery	=	80.280%	
35) Dibromofluoromethane	5.373	113	76768	47.858	ug/l	-0.01
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.720%	
50) Toluene-d8	8.641	98	276407	51.075	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.160%	
62) 4-Bromofluorobenzene	11.079	95	96274	50.912	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	101.820%	
Target Compounds						
17) Carbon Disulfide	2.508	76	1973	0.703	ug/l #	85
44) Trichloroethene	7.141	130	480	0.306	ug/l	94
64) Tetrachloroethene	9.275	164	314	0.225	ug/l #	67
68) m/p-Xylenes	10.305	106	763	0.267	ug/l	82
77) Bromobenzene	11.201	156	444	0.247	ug/l	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010625\
Data File : VX044584.D
Acq On : 06 Jan 2025 10:44
Operator : JC/MD
Sample : VX0106MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0106MBL01

Quant Time: Jan 06 16:02:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

