

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010625\
 Data File : VX044585.D
 Acq On : 06 Jan 2025 11:07
 Operator : JC/MD
 Sample : VX0106WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0106WBL01

Quant Time: Jan 06 16:02:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	138122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	249283	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	219263	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	101513	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	105902	43.730	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.460%
35) Dibromofluoromethane	5.379	113	85011	49.239	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.480%
50) Toluene-d8	8.646	98	298900	51.315	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.640%
62) 4-Bromofluorobenzene	11.079	95	105943	52.052	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.100%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010625\
Data File : VX044585.D
Acq On : 06 Jan 2025 11:07
Operator : JC/MD
Sample : VX0106WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0106WBL01

Quant Time: Jan 06 16:02:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

