

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010625\
 Data File : VX044588.D
 Acq On : 06 Jan 2025 12:23
 Operator : JC/MD
 Sample : Q1004-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190A-GW-803-805

Quant Time: Jan 06 16:03:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	129615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	224090	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	201326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	86406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96511	42.468	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.940%		
35) Dibromofluoromethane	5.385	113	80820	52.074	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	104.140%		
50) Toluene-d8	8.647	98	281592	53.779	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	107.560%		
62) 4-Bromofluorobenzene	11.079	95	93611	51.164	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	102.320%		
Target Compounds						
					Qvalue	
3) Chloromethane	1.294	50	705	0.368	ug/l #	51
16) Acetone	2.380	43	3879	4.187	ug/l	98
17) Carbon Disulfide	2.508	76	1568	0.568	ug/l #	83
20) Methylene Chloride	2.782	84	433	0.249	ug/l #	90
93) 1,2,4-Trichlorobenzene	13.591	180	584	0.366	ug/l	85
95) Naphthalene	13.780	128	3240	0.520	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010625\
Data File : VX044588.D
Acq On : 06 Jan 2025 12:23
Operator : JC/MD
Sample : Q1004-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190A-GW-803-805

Quant Time: Jan 06 16:03:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

