

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010625\
 Data File : VX044592.D
 Acq On : 06 Jan 2025 13:55
 Operator : JC/MD
 Sample : Q1004-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB190A-HYD-20241231

Quant Time: Jan 06 16:04:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	129626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	225020	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	199321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	88101	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	98751	43.450	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.900%		
35) Dibromofluoromethane	5.385	113	79024	50.706	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	101.420%		
50) Toluene-d8	8.647	98	266914	50.765	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.540%		
62) 4-Bromofluorobenzene	11.079	95	93598	50.946	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	101.900%		
Target Compounds						
3) Chloromethane	1.294	50	922	0.481	ug/l	94
16) Acetone	2.373	43	2256	2.435	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010625\
Data File : VX044592.D
Acq On : 06 Jan 2025 13:55
Operator : JC/MD
Sample : Q1004-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB190A-HYD-20241231

Quant Time: Jan 06 16:04:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

