

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010719\
 Data File : VX006939.D
 Acq On : 07 Jan 2019 16:48
 Operator : JC/MD
 Sample : K1012-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EO-02-010419-A

Quant Time: Jan 08 07:12:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	629089	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	957539	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	920116	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	446657	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	326366	48.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.28%	
35) Dibromofluoromethane	5.50	113	289700	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
50) Toluene-d8	8.71	98	1163991	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
62) 4-Bromofluorobenzene	11.13	95	415100	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
Target Compounds						
16) Acetone	2.45	43	40566	13.023	ug/l	100
18) Methyl Acetate	2.78	43	25278	2.760	ug/l	93
20) Methylene Chloride	2.86	84	21380	3.319	ug/l	98
43) Isopropyl Acetate	6.46	43	28349	2.257	ug/l	96
84) 1,2,4-Trimethylbenzene	11.80	105	27157	1.026	ug/l	96
95) Naphthalene	13.83	128	38014	1.159	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010719\
Data File : VX006939.D
Acq On : 07 Jan 2019 16:48
Operator : JC/MD
Sample : K1012-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
EO-02-010419-A

Quant Time: Jan 08 07:12:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

