


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010719\
 Data File : VX006924.D
 Acq On : 07 Jan 2019 09:47
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Title : SW846 8260
 Last Update : Wed Dec 19 15:20:58 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1628

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	20679	PASS
75	95	30	60	46.8	56066	PASS
95	95	100	100	100.0	119778	PASS
96	95	5	9	6.6	7911	PASS
173	174	0.00	2	0.7	785	PASS
174	95	50	100	89.8	107533	PASS
175	174	5	9	7.4	7932	PASS
176	174	95	101	96.7	103936	PASS
177	176	5	9	6.4	6611	PASS