

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010720\
 Data File : VX014455.D
 Acq On : 07 Jan 2020 18:08
 Operator : JC/SP
 Sample : K6347-02 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Jan 08 03:04:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	553245	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	831159	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	766803	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	386148	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	291651	47.22	ug/l	0.00
Spiked Amount			Recovery	=	94.44%	
35) Dibromofluoromethane	5.49	113	247987	49.29	ug/l	0.00
Spiked Amount			Recovery	=	98.58%	
50) Toluene-d8	8.71	98	998820	51.16	ug/l	0.00
Spiked Amount			Recovery	=	102.32%	
62) 4-Bromofluorobenzene	11.13	95	356076	49.68	ug/l	0.00
Spiked Amount			Recovery	=	99.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.02	59	103398	80.181	ug/l	97
31) Cyclohexane	5.58	56	93630	9.918	ug/l	99
39) Methylcyclohexane	7.46	83	32070	3.266	ug/l	96
40) Benzene	6.14	78	3200724	137.971	ug/l	100
42) 1,2-Dichloroethane	6.18	62	64617	7.966	ug/l	99
52) Toluene	8.78	92	5204	0.350	ug/l	97
68) m/p-Xylenes	10.35	106	4032	0.368	ug/l	77
69) o-Xylene	10.70	106	193152	18.500	ug/l	99
73) Isopropylbenzene	11.01	105	17107	0.631	ug/l	92
89) n-Butylbenzene	12.39	91	10361	0.462	ug/l	95
91) 1,2-Dichlorobenzene	12.39	146	3965	0.300	ug/l	85
95) Naphthalene	13.83	128	39304	1.365	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010720\
Data File : VX014455.D
Acq On : 07 Jan 2020 18:08
Operator : JC/SP
Sample : K6347-02 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Jan 08 03:04:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

