

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010721\
 Data File : VX020527.D
 Acq On : 07 Jan 2021 12:28
 Operator : JC/MD
 Sample : VX0107MBL01
 Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK401

Quant Time: Jan 08 00:56:21 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 08 00:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.830	114	333392	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.092	117	318418	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.061	152	134976	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.386	65	133554	49.86	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	99.72%	
7) Chloroethane-d5	1.697	69	108963	52.05	ug/L	0.01
Spiked Amount	50.000	Range 70 - 130	Recovery	=	104.10%	
11) 1,1-Dichloroethene-d2	2.343	63	189502	37.88	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	75.76%	
21) 2-Butanone-d5	4.532	46	180914	110.23	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	110.23%	
24) Chloroform-d	5.135	84	255782	55.50	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	111.00%	
26) 1,2-Dichloroethane-d4	6.025	65	166069	54.41	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	108.82%	
32) Benzene-d6	6.044	84	519927	54.28	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	108.56%	
36) 1,2-Dichloropropane-d6	7.367	67	163882	55.33	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	110.66%	
41) Toluene-d8	8.696	98	461154	53.11	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.22%	
43) trans-1,3-Dichloroprop...	8.988	79	77185	53.57	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	107.14%	
47) 2-Hexanone-d5	9.427	63	137661	105.23	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	105.23%	
56) 1,1,2,2-Tetrachloroeth...	11.232	84	211624	51.44	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	102.88%	
66) 1,2-Dichlorobenzene-d4	12.360	152	155303	58.70	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	117.40%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010721\
Data File : VX020527.D
Acq On : 07 Jan 2021 12:28
Operator : JC/MD
Sample : VX0107MBL01
Misc : 5.0g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK401

Quant Time: Jan 08 00:56:21 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 08 00:55:52 2021
Response via : Initial Calibration

