

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010721\
 Data File : VX020528.D
 Acq On : 07 Jan 2021 13:59
 Operator : JC/MD
 Sample : VX0107WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK402

Quant Time: Jan 08 00:56:27 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Jan 08 00:55:52 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.824	114	335633	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.098	117	314170	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.061	152	133455	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.386	65	138309	51.29	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	102.58%
7) Chloroethane-d5	1.691	69	110749	52.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.10%
11) 1,1-Dichloroethene-d2	2.343	63	195328	38.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.56%
21) 2-Butanone-d5	4.526	46	175171	106.02	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.02%
24) Chloroform-d	5.135	84	259147	55.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.70%
26) 1,2-Dichloroethane-d4	6.025	65	165811	53.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.94%
32) Benzene-d6	6.044	84	528639	55.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	111.88%
36) 1,2-Dichloropropane-d6	7.367	67	164505	56.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.58%
41) Toluene-d8	8.696	98	439181	51.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.52%
43) trans-1,3-Dichloroprop...	8.988	79	69851	49.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.26%
47) 2-Hexanone-d5	9.421	63	130508	101.11	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	101.11%
56) 1,1,2,2-Tetrachloroeth...	11.232	84	216994	53.46	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.92%
66) 1,2-Dichlorobenzene-d4	12.360	152	154858	59.20	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	118.40%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010721\
Data File : VX020528.D
Acq On : 07 Jan 2021 13:59
Operator : JC/MD
Sample : VX0107WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK402

Quant Time: Jan 08 00:56:27 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Jan 08 00:55:52 2021
Response via : Initial Calibration

