

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010725\
 Data File : VX044608.D
 Acq On : 07 Jan 2025 17:42
 Operator : JC/MD
 Sample : Q1004-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190A-GW-778-780

Quant Time: Jan 08 05:56:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	172015	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	313643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	283361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	116507	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	138153	49.811	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.620%		
35) Dibromofluoromethane	5.379	113	100812	46.874	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.740%		
50) Toluene-d8	8.647	98	371289	50.541	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	101.080%		
62) 4-Bromofluorobenzene	11.079	95	130449	50.309	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	100.620%		
Target Compounds						
16) Acetone	2.380	43	3556	4.068	ug/l	94
44) Trichloroethene	7.129	130	9851	4.401	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010725\
Data File : VX044608.D
Acq On : 07 Jan 2025 17:42
Operator : JC/MD
Sample : Q1004-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190A-GW-778-780

Quant Time: Jan 08 05:56:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 08 03:57:12 2025
Response via : Initial Calibration

