

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010725\
 Data File : VX044610.D
 Acq On : 07 Jan 2025 18:28
 Operator : JC/MD
 Sample : Q1022-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190A-GW-838-840

Quant Time: Jan 08 05:56:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 08 03:57:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	188583	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	335673	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	308993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	131138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	153060	50.337	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 100.680%		
35) Dibromofluoromethane	5.379	113	109447	47.549	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.100%		
50) Toluene-d8	8.647	98	403684	51.344	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.680%		
62) 4-Bromofluorobenzene	11.079	95	143530	51.721	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 103.440%		
Target Compounds						
16) Acetone	2.374	43	6058	6.321	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010725\
Data File : VX044610.D
Acq On : 07 Jan 2025 18:28
Operator : JC/MD
Sample : Q1022-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190A-GW-838-840

Quant Time: Jan 08 05:56:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X010725W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 08 03:57:12 2025
Response via : Initial Calibration

