


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010819\
Data File : VX006946.D
Acq On : 08 Jan 2019 09:44
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010819W.M
Title : SW846 8260
Last Update : Tue Jan 08 14:06:59 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	20083	PASS
75	95	30	60	47.1	55669	PASS
95	95	100	100	100.0	118074	PASS
96	95	5	9	6.9	8132	PASS
173	174	0.00	2	1.3	1337	PASS
174	95	50	100	87.4	103216	PASS
175	174	5	9	7.4	7671	PASS
176	174	95	101	99.8	102960	PASS
177	176	5	9	6.5	6713	PASS