

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
 Data File : VX014467.D
 Acq On : 08 Jan 2020 13:38
 Operator : JC/SP
 Sample : L1034-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-C

Quant Time: Jan 08 15:49:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	521363	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	806029	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	730732	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	333470	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	287501	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
35) Dibromofluoromethane	5.49	113	236614	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
50) Toluene-d8	8.71	98	955245	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.14	95	326853	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
Target Compounds						
19) Methyl tert-butyl Ether	3.19	73	15565	0.942	ug/l #	Qvalue 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
Data File : VX014467.D
Acq On : 08 Jan 2020 13:38
Operator : JC/SP
Sample : L1034-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-C

Quant Time: Jan 08 15:49:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

