

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
 Data File : VX014469.D
 Acq On : 08 Jan 2020 14:24
 Operator : JC/SP
 Sample : PB125936TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125936TB

Quant Time: Jan 08 15:55:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	541045	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	833095	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	755155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	345883	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	296135	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
35) Dibromofluoromethane	5.49	113	247370	49.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.10%	
50) Toluene-d8	8.71	98	982900	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	11.14	95	344260	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
Target Compounds						
16) Acetone	2.43	43	21130	2.846	ug/l	89
20) Methylene Chloride	2.86	84	11855	1.951	ug/l	92
43) Isopropyl Acetate	6.43	43	137700	10.017	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	12619	0.626	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
Data File : VX014469.D
Acq On : 08 Jan 2020 14:24
Operator : JC/SP
Sample : PB125936TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125936TB

Quant Time: Jan 08 15:55:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

