

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
 Data File : VX014473.D
 Acq On : 08 Jan 2020 15:58
 Operator : JC/SP
 Sample : K6474-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-01

Quant Time: Jan 09 06:53:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	522640	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	817038	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	742515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	341887	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	286668	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
35) Dibromofluoromethane	5.49	113	239149	48.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.70%	
50) Toluene-d8	8.71	98	969928	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
62) 4-Bromofluorobenzene	11.14	95	332759	47.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.46%	
Target Compounds						
20) Methylene Chloride	2.85	84	12207	2.080	ug/l	95
43) Isopropyl Acetate	6.43	43	131147	9.727	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
Data File : VX014473.D
Acq On : 08 Jan 2020 15:58
Operator : JC/SP
Sample : K6474-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-01

Quant Time: Jan 09 06:53:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

