

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
 Data File : VX014474.D
 Acq On : 08 Jan 2020 16:21
 Operator : JC/SP
 Sample : K6474-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-02

Quant Time: Jan 09 06:57:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	539598	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	816511	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	748563	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	345207	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	291708	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
35) Dibromofluoromethane	5.49	113	243276	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	8.71	98	979390	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
62) 4-Bromofluorobenzene	11.13	95	339465	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
Target Compounds						
20) Methylene Chloride	2.85	84	10523	1.737	ug/l	90
43) Isopropyl Acetate	6.43	43	129132	9.584	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
Data File : VX014474.D
Acq On : 08 Jan 2020 16:21
Operator : JC/SP
Sample : K6474-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-02

Quant Time: Jan 09 06:57:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

