

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
 Data File : VX014481.D
 Acq On : 08 Jan 2020 19:03
 Operator : JC/SP
 Sample : PB125936ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125936ZHE#05

Quant Time: Jan 09 07:37:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	526703	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	822659	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	737698	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	342001	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	290409	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
35) Dibromofluoromethane	5.49	113	240817	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
50) Toluene-d8	8.71	98	975371	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	11.13	95	337469	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
Target Compounds						
20) Methylene Chloride	2.85	84	12508	2.115	ug/l	94
43) Isopropyl Acetate	6.43	43	137783	10.150	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
Data File : VX014481.D
Acq On : 08 Jan 2020 19:03
Operator : JC/SP
Sample : PB125936ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125936ZHE#05

Quant Time: Jan 09 07:37:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

