

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
 Data File : VX014482.D
 Acq On : 08 Jan 2020 19:27
 Operator : JC/SP
 Sample : PB125936ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125936ZHE#06

Quant Time: Jan 09 07:50:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	521808	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	805500	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	747846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	345183	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	287785	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
35) Dibromofluoromethane	5.49	113	239370	49.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.18%	
50) Toluene-d8	8.71	98	967296	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
62) 4-Bromofluorobenzene	11.13	95	338494	48.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.46%	
Target Compounds						
20) Methylene Chloride	2.85	84	12772	2.179	ug/l	95
43) Isopropyl Acetate	6.43	43	139129	10.467	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX010820\
Data File : VX014482.D
Acq On : 08 Jan 2020 19:27
Operator : JC/SP
Sample : PB125936ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125936ZHE#06

Quant Time: Jan 09 07:50:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.65 min Scan# 782

Delta R.T. 0.00 min

Lab File: VX014482.D

Acq: 08 Jan 2020 19:27

Instrument :

MSVOA_X

ClientSampleId :


PB125936ZHE#06

Tgt Ion:168 Resp: 521808

Ion Ratio Lower Upper




168 100

99 50.9 39.2 58.8



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#20

Methylene Chloride

Concen: 2.179 ug/l

RT: 2.85 min Scan# 322

Delta R.T. 0.00 min

Lab File: VX014482.D

Acq: 08 Jan 2020 19:27

Tgt Ion: 84 Resp: 12772


Ion Ratio Lower Upper

84 100

49 117.3 95.8 143.8

51 35.9 30.2 45.4

86 55.0 51.8 77.6






Abundance Ion 83.90 (83.60 to 84.60): VX0

Ion 48.90 (48.60 to 49.60): VX0

Ion 50.90 (50.60 to 51.60): VX0

Ion 85.90 (85.60 to 86.60): VX0









#33
 1,2-Dichloroethane-d4
 Concen: 49.404 ug/l
 RT: 6.05 min Scan# 847
 Delta R.T. 0.00 min
 Lab File: VX014482.D
 Acq: 08 Jan 2020 19:27




Instrument :
 MSVOA_X
 ClientSampleId :
 PB125936ZHE#06


Tgt Ion: 65 Resp: 287785
 Ion Ratio Lower Upper
 65 100
 67 52.7 0.0 106.2



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.85 min Scan# 978
 Delta R.T. 0.00 min
 Lab File: VX014482.D
 Acq: 08 Jan 2020 19:27

Tgt Ion: 114 Resp: 805500
 Ion Ratio Lower Upper
 114 100
 63 20.4 0.0 39.6
 88 15.9 0.0 30.8









#35
 Dibromofluoromethane
 Concen: 49.091 ug/l
 RT: 5.49 min Scan# 755
 Delta R.T. 0.00 min
 Lab File: VX014482.D
 Acq: 08 Jan 2020 19:27




Instrument :
 MSVOA_X
 ClientSampleId :
 PB125936ZHE#06


Tgt Ion	Ratio	Lower	Upper
113	100		
111	103.5	81.5	122.3
192	23.5	19.2	28.8



#43
 Isopropyl Acetate
 Concen: 10.467 ug/l
 RT: 6.43 min Scan# 910
 Delta R.T. 0.00 min
 Lab File: VX014482.D
 Acq: 08 Jan 2020 19:27

Tgt Ion	Ratio	Lower	Upper
43	100		
61	20.4	16.6	24.8
87	13.3	10.8	16.2





#50

Toluene-d8

Concen: 51.119 ug/l

RT: 8.71 min Scan# 1283

Delta R.T. 0.00 min

Lab File: VX014482.D

Acq: 08 Jan 2020 19:27

Instrument :

MSVOA_X

ClientSampleId :





PB125936ZHE#06

Tgt Ion: 98 Resp: 967296

Ion Ratio Lower Upper

98 100

100 65.7 53.2 79.8



#62

4-Bromofluorobenzene

Concen: 48.728 ug/l

RT: 11.13 min Scan# 1680

Delta R.T. 0.00 min

Lab File: VX014482.D

Acq: 08 Jan 2020 19:27




Tgt Ion: 95 Resp: 338494


Ion Ratio Lower Upper

95 100

174 89.2 0.0 180.8

176 86.8 0.0 172.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX014482.D
Acq: 08 Jan 2020 19:27

Instrument :
MSVOA_X
ClientSampleId :
PB125936ZHE#06

Tgt Ion:117 Resp: 747846
Ion Ratio Lower Upper
117 100
82 51.5 41.8 62.6
119 32.7 26.2 39.4




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX014482.D
Acq: 08 Jan 2020 19:27

Tgt Ion:152 Resp: 345183
Ion Ratio Lower Upper
152 100
115 55.8 39.5 118.3
150 156.6 0.0 350.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

