


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010820\
 Data File : VX014459.D
 Acq On : 08 Jan 2020 09:15
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Title : SW846 8260
 Last Update : Tue Jan 07 15:54:41 2020



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	26956	PASS
75	95	30	60	49.7	69058	PASS
95	95	100	100	100.0	138912	PASS
96	95	5	9	6.8	9429	PASS
173	174	0.00	2	0.8	906	PASS
174	95	50	100	85.0	118099	PASS
175	174	5	9	7.6	8950	PASS
176	174	95	101	97.9	115629	PASS
177	176	5	9	6.3	7239	PASS