

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010821\
 Data File : VX020535.D
 Acq On : 08 Jan 2021 11:03
 Operator : JC/MD
 Sample : VX0108WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK404

Quant Time: Jan 09 04:42:07 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jan 09 04:41:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.824	114	307386	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.092	117	302003	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.061	152	126444	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.386	65	120656	48.86	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	97.72%
7) Chloroethane-d5	1.691	69	101406	52.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.08%
11) 1,1-Dichloroethene-d2	2.343	63	180148	39.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.12%
21) 2-Butanone-d5	4.526	46	177324	117.18	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	117.18%
24) Chloroform-d	5.135	84	245202	57.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	115.40%
26) 1,2-Dichloroethane-d4	6.025	65	161621	57.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	114.88%
32) Benzene-d6	6.044	84	493974	54.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.76%
36) 1,2-Dichloropropane-d6	7.366	67	158311	56.36	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	112.72%
41) Toluene-d8	8.695	98	431445	52.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	104.78%
43) trans-1,3-Dichloroprop...	8.988	79	73732	53.95	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	107.90%
47) 2-Hexanone-d5	9.421	63	139635	112.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	112.54%
56) 1,1,2,2-Tetrachloroeth...	11.232	84	206937	53.04	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.08%
66) 1,2-Dichlorobenzene-d4	12.359	152	146187	58.98	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	117.96%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010821\
Data File : VX020535.D
Acq On : 08 Jan 2021 11:03
Operator : JC/MD
Sample : VX0108WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK404

Quant Time: Jan 09 04:42:07 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jan 09 04:41:43 2021
Response via : Initial Calibration

