

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010821\
 Data File : VX020538.D
 Acq On : 08 Jan 2021 14:49
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Jan 09 04:43:01 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jan 09 04:41:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.824	114	217560	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.092	117	216044	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.061	152	91666	50.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.386	65	118933	68.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	136.08%#
7) Chloroethane-d5	1.691	69	99140	72.57	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	145.14%#
11) 1,1-Dichloroethene-d2	2.343	63	171341	52.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	104.96%
21) 2-Butanone-d5	4.532	46	171292	159.94	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	159.94%#
24) Chloroform-d	5.135	84	239939	79.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	159.56%#
26) 1,2-Dichloroethane-d4	6.025	65	155019	77.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	155.66%#
32) Benzene-d6	6.044	84	480903	74.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	148.00%#
36) 1,2-Dichloropropane-d6	7.367	67	152799	76.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	152.08%#
41) Toluene-d8	8.696	98	404563	68.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	137.34%#
43) trans-1,3-Dichloroprop...	8.994	79	67235	68.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	137.54%#
47) 2-Hexanone-d5	9.421	63	131802	148.50	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	148.50%#
56) 1,1,2,2-Tetrachloroeth...	11.232	84	203200	72.80	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	145.60%#
66) 1,2-Dichlorobenzene-d4	12.360	152	146667	81.62	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	163.24%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010821\
Data File : VX020538.D
Acq On : 08 Jan 2021 14:49
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Jan 09 04:43:01 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jan 09 04:41:43 2021
Response via : Initial Calibration

