

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010821\
 Data File : VX020540.D
 Acq On : 08 Jan 2021 15:47
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Jan 09 04:43:24 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Jan 09 04:41:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.824	114	349791	50.00	ug/L	0.00
28) Chlorobenzene-d5	10.092	117	332772	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.061	152	144034	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.386	65	105710	37.61	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.22%
7) Chloroethane-d5	1.691	69	90733	41.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.62%
11) 1,1-Dichloroethene-d2	2.343	63	152652	29.08	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.16%#
21) 2-Butanone-d5	4.526	46	153217	88.98	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	88.98%
24) Chloroform-d	5.135	84	211499	43.74	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.48%
26) 1,2-Dichloroethane-d4	6.025	65	136897	42.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.50%
32) Benzene-d6	6.044	84	427723	42.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.46%
36) 1,2-Dichloropropane-d6	7.367	67	135591	43.81	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.62%
41) Toluene-d8	8.689	98	356522	39.29	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.58%#
43) trans-1,3-Dichloroprop...	8.988	79	59978	39.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.66%
47) 2-Hexanone-d5	9.421	63	110166	80.58	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.58%
56) 1,1,2,2-Tetrachloroeth...	11.232	84	173017	40.24	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.48%
66) 1,2-Dichlorobenzene-d4	12.360	152	125503	44.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.90%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010821\
Data File : VX020540.D
Acq On : 08 Jan 2021 15:47
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Jan 09 04:43:24 2021
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SFAMXML010721WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Jan 09 04:41:43 2021
Response via : Initial Calibration

