

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010824\
 Data File : VX039659.D
 Acq On : 08 Jan 2024 12:26
 Operator : JC/MD
 Sample : P1015-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-301P-25-27

Quant Time: Jan 09 04:33:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	107173	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	209910	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92121	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85881	43.742	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	87.480%		
35) Dibromofluoromethane	5.385	113	67105	46.274	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.540%		
50) Toluene-d8	8.647	98	262391	48.738	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.480%		
62) 4-Bromofluorobenzene	11.079	95	108546	48.866	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.740%		
Target Compounds						
20) Methylene Chloride	2.788	84	7697	4.703	ug/l	95
25) 2-Butanone	4.562	43	22575	13.325	ug/l	98
43) Isopropyl Acetate	6.342	43	91264	19.851	ug/l	99
95) Naphthalene	13.780	128	7697	1.142	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010824\
Data File : VX039659.D
Acq On : 08 Jan 2024 12:26
Operator : JC/MD
Sample : P1015-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-301P-25-27

Quant Time: Jan 09 04:33:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

