

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010824\
 Data File : VX039660.D
 Acq On : 08 Jan 2024 12:49
 Operator : JC/MD
 Sample : P1056-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-311P-0-2

Quant Time: Jan 09 04:33:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	103802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211468	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87986	46.270	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.540%
35) Dibromofluoromethane	5.385	113	67601	46.273	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.540%
50) Toluene-d8	8.647	98	262525	48.403	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.800%
62) 4-Bromofluorobenzene	11.079	95	109569	48.963	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.920%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	44258	18.389	ug/l	92
20) Methylene Chloride	2.788	84	5252	3.313	ug/l	91
43) Isopropyl Acetate	6.342	43	95229	20.561	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010824\
Data File : VX039660.D
Acq On : 08 Jan 2024 12:49
Operator : JC/MD
Sample : P1056-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-311P-0-2

Quant Time: Jan 09 04:33:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

