

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010824\
 Data File : VX039661.D
 Acq On : 08 Jan 2024 13:12
 Operator : JC/MD
 Sample : P1056-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MR-311P-30-32

Quant Time: Jan 09 04:34:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	101512	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	204996	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207273	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91102	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86030	46.262	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.520%
35) Dibromofluoromethane	5.379	113	67020	47.324	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.640%
50) Toluene-d8	8.647	98	257320	48.941	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.880%
62) 4-Bromofluorobenzene	11.079	95	108560	50.043	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.080%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	59242	34.623	ug/l	92
20) Methylene Chloride	2.788	84	5015	3.235	ug/l	92
43) Isopropyl Acetate	6.336	43	42967	9.570	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010824\
Data File : VX039661.D
Acq On : 08 Jan 2024 13:12
Operator : JC/MD
Sample : P1056-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MR-311P-30-32

Quant Time: Jan 09 04:34:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

