

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010824\
 Data File : VX039662.D
 Acq On : 08 Jan 2024 13:34
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 09 04:34:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	98561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	198847	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87157	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	81299	45.027	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.060%
35) Dibromofluoromethane	5.385	113	63812	46.452	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.900%
50) Toluene-d8	8.647	98	245753	48.187	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.380%
62) 4-Bromofluorobenzene	11.079	95	100111	47.576	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010824\
Data File : VX039662.D
Acq On : 08 Jan 2024 13:34
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 09 04:34:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

