

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010920\
 Data File : VX014493.D
 Acq On : 09 Jan 2020 12:53
 Operator : JC/SP
 Sample : L1044-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-WILLETS-PT-BLVD-(JAN)

Quant Time: Jan 09 15:22:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X121219W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 13 01:35:34 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.01	128	79215	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.85	114	438700	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.11	117	392338	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.05	65	168792	29.75	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	99.17%
60) 4-Bromofluorobenzene	11.13	95	181900	28.71	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	95.70%
63) Toluene-d8	8.71	98	522643	29.80	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.33%

Target Compounds

					Ovalue
12) Methyl Acetate	2.76	43	18341	2.671	ug/l 93
15) Acetone	2.43	58	1037476	1315.553	ug/l 92
30) 2-Butanone	4.66	43	31519	8.439	ug/l # 77
58) 4-Methyl-2-Pentanone	8.63	43	17066	2.325	ug/l 97
62) Toluene	8.78	91	147724	7.079	ug/l 98
80) 1,2,4-Trimethylbenzene	11.81	105	6705	0.354	ug/l 50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010920\
Data File : VX014493.D
Acq On : 09 Jan 2020 12:53
Operator : JC/SP
Sample : L1044-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-WILLETS-PT-BLVD-(JAN)

Quant Time: Jan 09 15:22:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X121219W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 13 01:35:34 2019
Response via : Initial Calibration

