

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010920\
 Data File : VX014507.D
 Acq On : 09 Jan 2020 19:20
 Operator : JC/SP
 Sample : PB125963TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125963TB

Quant Time: Jan 10 07:33:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	508166	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	789809	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	726491	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	337249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	279052	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
35) Dibromofluoromethane	5.49	113	235161	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
50) Toluene-d8	8.71	98	941086	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
62) 4-Bromofluorobenzene	11.13	95	325693	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
Target Compounds						
20) Methylene Chloride	2.86	84	8999	1.577	ug/l	84
43) Isopropyl Acetate	6.43	43	139646	10.715	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010920\
Data File : VX014507.D
Acq On : 09 Jan 2020 19:20
Operator : JC/SP
Sample : PB125963TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125963TB

Quant Time: Jan 10 07:33:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

