

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010920\
 Data File : VX014508.D
 Acq On : 09 Jan 2020 19:43
 Operator : JC/SP
 Sample : K6456-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-03

Quant Time: Jan 10 07:37:05 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	529568	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	813977	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	744242	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	329705	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	291169	49.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.50%	
35) Dibromofluoromethane	5.49	113	242937	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
50) Toluene-d8	8.71	98	980145	51.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.52%	
62) 4-Bromofluorobenzene	11.13	95	330962	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
Target Compounds						
20) Methylene Chloride	2.85	84	10597	1.782	ug/l	96
43) Isopropyl Acetate	6.43	43	119703	8.912	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010920\
Data File : VX014508.D
Acq On : 09 Jan 2020 19:43
Operator : JC/SP
Sample : K6456-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-03

Quant Time: Jan 10 07:37:05 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

