

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX010920\
 Data File : VX014517.D
 Acq On : 09 Jan 2020 23:12
 Operator : JC/SP
 Sample : PB125964ZHE#10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB125964ZHE#10

Quant Time: Jan 10 08:13:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 07 15:54:41 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	500909	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	779865	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	714889	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	333335	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	277158	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
35) Dibromofluoromethane	5.49	113	231997	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
50) Toluene-d8	8.71	98	932627	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
62) 4-Bromofluorobenzene	11.13	95	323557	48.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.22%	
Target Compounds						
20) Methylene Chloride	2.85	84	10372	1.844	ug/l	93
43) Isopropyl Acetate	6.43	43	137597	10.692	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX010920\
Data File : VX014517.D
Acq On : 09 Jan 2020 23:12
Operator : JC/SP
Sample : PB125964ZHE#10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB125964ZHE#10

Quant Time: Jan 10 08:13:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X010720W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 07 15:54:41 2020
Response via : Initial Calibration

