

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010923\
 Data File : VX033665.D
 Acq On : 09 Jan 2023 16:02
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jan 10 00:47:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jan 07 06:16:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	12150	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	78583	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	81295	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	33210	33.757	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	112.533%#
60) 4-Bromofluorobenzene	11.079	95	43008	29.353	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	97.833%
63) Toluene-d8	8.647	98	80320	30.872	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	102.900%
Target Compounds						
					Qvalue	
15) Acetone	2.392	58	236	1.627	ug/l	70
18) Methylene Chloride	2.788	84	2060	1.585	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010923\
Data File : VX033665.D
Acq On : 09 Jan 2023 16:02
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jan 10 00:47:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jan 07 06:16:21 2023
Response via : Initial Calibration

