

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010923\
 Data File : VX033667.D
 Acq On : 09 Jan 2023 16:49
 Operator : JC/MD
 Sample : 01079-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SAMPLE-2-INFLUENT

Quant Time: Jan 10 00:47:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jan 07 06:16:21 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/10/2023
 Supervised By :Mahesh Dadoda 01/11/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	8760m	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	84313	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	91347	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	32961	46.470	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	154.900%#
60) 4-Bromofluorobenzene	11.079	95	30888	18.761	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	62.533%#
63) Toluene-d8	8.647	98	87215	29.833	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.433%
Target Compounds						
15) Acetone	2.386	58	279m	2.668	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX010923\
Data File : VX033667.D
Acq On : 09 Jan 2023 16:49
Operator : JC/MD
Sample : 01079-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SAMPLE-2-INFLUENT

Quant Time: Jan 10 00:47:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X010723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jan 07 06:16:21 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/10/2023
Supervised By :Mahesh Dadoda 01/11/2023

